# Can Legflow improve treatment of long femoropopliteal lesions: The REFLOW outcomes

Dr. Michel Bosiers CX 2021




# My disclosures


- X I do not have any potential conflicts of interest to report
- I have the following potential conflicts of interest to report:
  - ☐ Consulting
  - ☐ Employment in industry
  - ☐ Stockholder of a healthcare company
  - ☐ Owner of a healthcare company
  - □Other(s)

#### **DCB**


# DCB-treatment works... Proof of concepts



# 1-Year Patency Rates of DCB (in ideal circumstances)



# However in "Real Life"...





# REFLOW study



A study investigating the Efficacy of the LEGFLOW Paclitaxel-Eluting for the treatment of long femoropopliteal lesions(TASC C&D)





**SAFEPAX®** matrix system, proprietary formulation of Lipophilic and Polymeric Ammonium Salt based excipient for PTX DCBs

Minimum wash off rate

Prevents PTX crystallization on the balloons surface

CARDIONOVUM portfolio

#### **LEGFLOW® RX/OTW**

Peripheral Balloons Dilatation Range. Platforms: 0.014" 0.018" 0.035"

#### **APERTO**° OTW

Hemodialysis Shunt Balloon, up to 20 bar Platform: 0.0.035"

#### **RESTORE®** DEB

Paclitaxel Releasing PTCA Balloon Catheter Platform: 0.014"

Drug:

PTX (3.0μg/mm²)

Excipient:

Ammonium Salt

Prevents crystallization

Amorphus

Coating Technology

Restenosis

prevention

Prevention

Homogeneous coating
and minimal mass effect





**OLD: CRYSTALLINE** 

**NEW: AMORPHOUS** 

**Coating look** 



Sugar cube (crystalline sugar).



Honey (non-crystalline sugar)

Matrix



Crystalline

Crystalline excipients (ex: Hurea, Sorbitol, etc.) and crystalline PTX: they built a rigid crystals shape (white crust formed by crystals of PTX +excipient)



**Amorphous** 

Polymeric excipients (ex: Ammonium salt) and noncrystalline PTX: "melt" PTX in an elastic matrix.

A Polymeric excipient is made by very long molecules chain without a specific and rigid shape

#### **COATING CLASSES**


#### **OLD: CRYSTALLINE**

**NEW: AMORPHOUS** 

Mechanical stress response



Crystalline coating is affected by mechanical stress due to the rigid structure of <u>crystralline</u> excipients.



Amorphous coating is not affected by mechanical stress due to the elastic, polymeric excipients



During inflation, early DCBs generations coating is affected by flaking.



During inflation, Safepax (Amorphous coating) is not affected by flaking.
Ammonium Salt prevents any crystallization on the balloon surface.

Cardionovum has the highest coating stability (paclitaxel loss during handling & wash-off) of DCBs without compromising performance

**CARDIONOVUM** 

COMPETITORS

# Study design



#### Study Objective:

To evaluate the performance of the **LEGFLOW Paclitaxel-Eluting** Peripheral balloon catheter for the treatment of **long femoropopliteal lesions (TASC C&D)**.

### Primary Endpoint:

Primary Patency at 12 months, defined as absence of a hemodynamically significant stenosis on duplex ultrasound (systolic velocity ratio ≤2.4) at the target lesion and without reintervention.

# Participating centers



#### BELGIUM

- M. Bosiers, K. Deloose, J. Callaert AZ Sint-Blasius, Dendermonde
- P. Peeters, J. Verbist, W. Van den Eynde Imelda Hospital, Bonheiden
- L. Maene, R. Beelen OLV, Aalst
- K. Keirse RZ Heilig Hart, Tienen
- J. Hendriks, P. Lauwers University Hospital Antwerp, Edegem

#### GERMANY

- G. Torsello St. Franziskus-Hospital Münster
- D. Scheinert Universitätsklinikum Leipzig

# Inclusion criteria



# 120 out of 120 patients enrolled (100%)

#### Main inclusion criteria

- Rutherford classification from 2 to 5
- **De novo lesion** in the femoropopliteal arteries, suitable for endovascular therapy
- Total target lesion length > 150mm

# Study overview




Timeline	Baseline	Disch	1M	6M	12M	24M
Medication  Physical examination						
Rutherford						
ABI Core Lab Ultrasound						
Color Flow Ultrasou	nd					

# Patient Demographics



	N = 120
Male (%)	65.80% (79/120)
Age (min – max)	<b>71.06</b> (35.05 – 93.16) years
Nicotine abuse (%)	56.67% (68/120)
Hypertension (%)	<b>77.50</b> % (93/120)
Diabetes mellitus (%)	30.00% (36/120)
Renal insufficiency (%)	15.00% (18/120)
Hypercholesterolemia (%)	53.30% (64/120)
Obesity (%)	19.20% (23/120)



# Procedural characteristics



	N = 120
Procedure time (min-max)	<b>52.17</b> (19-165) minutes
Scopy time (min – max)	7.32 (1.7 – 39.24) minutes *missing information for 2 patients
Contrast (min – max)	88.09 (9 – 195) mL
Cross-over performed (%)	83.33% (100/120)
Inflow Lesion (%)	10.83% (13/120)
Outflow lesion (%)	21.67% (26/120)

# Lesion Characteristics



	N = 120
Lesion length (min – max)	<b>216.08</b> (150 – 390) mm
Ref Vessel Diameter (min – max)	5.40 (4.05 – 6.00) mm
Pre-dilatation	64.20% (77/120)
1 DCB (%) 2 DCB's (%) 3 DCB's (%)	25.83% (31/120) 57.50% (69/120) 16.67% (20/120)
Post-dilatation (%)	22.50% (27/120)
Bail-out stenting (%)	35.00% (42/120)
Occlusion (%)	45.00% (54/120)
Calcified lesion (%)	67.50% (81/120)

# Paclitaxel --> mortality?

# Risk of Death Following Application of Paclitaxel Stents in the Femoropopliteal Artery of the Leg: Meta-Analysis of Randomized Controlled Trials

Konstantinos Katsanos ☑, Stavros Spiliopoulos, Panagiotis Kitrou, Miltiadis Krokidis, anc • Originally published 6 Dec 2018 | https://doi.org/10.1161/JAHA.118.011245 | Journal of the American H€

#### Study

ZILVER-PTX <sup>19</sup>
FINN-PTX <sup>18</sup>
IN.PACT SFA <sup>82</sup>
FEMPAC <sup>29</sup>
LEVANT II <sup>27</sup>
LEVANT II <sup>26</sup>
CONSEQUENT <sup>30</sup>
ILLUMENATE EU <sup>32</sup>
ISAR-STATH <sup>51</sup>
ISAR-PEBIS <sup>55</sup>
ACOART I <sup>40</sup>
IN.PACT SFA JAPAN <sup>41</sup>

Fixed effect model Random effects model

Heterogeneity:  $I^2 = 0\%$ ,  $\tau^2 = 0$ , p = 0.80

#### Recommendations

Based on the FDA's review of available data and the advisory panel conclusions, the FDA recommends that health care providers consider the following:

- Continue diligent monitoring of patients who have been treated with paclitaxel-coated balloons and paclitaxel-eluting stents.
- When making treatment recommendations, and as part of the informed consent process, consider that there may be an increased rate of long-term mortality in patients treated with paclitaxel-coated balloons and paclitaxel-eluting stents.

•	Discuss the risks and benefits of all available PAD treatment options with patients. For
	many patients, alternative treatment options to paclitaxel-coated balloons and
	paclitaxel-eluting stents provide a more favorable benefit-risk profile based on
	currently available information.

- For individual patients judged to be at particularly high risk for restenosis and repeat femoropopliteal interventions, clinicians may determine that the benefits of using a paclitaxel-coated device outweigh the risk of late mortality.
- In discussing treatment options, physicians should explore their patients' expectations, concerns, and treatment preferences.
- Ensure patients receive optimal medical therapy for PAD and other cardiovascular risk factors as well as guidance on healthy lifestyles including weight control, smoking cessation, and exercise.
- Report any adverse events or suspected adverse events experienced with the use of paclitaxel-coated balloons and paclitaxel-eluting stents.

0.01 0.1 1 10 100



Veight

ndom)

20.6%

1.5%

3.7%

8.9%

9.4%

21.4%

2.6%

9.9%

2.9%

1.7%


14.2%

00.0%

3.2%

# 24-month Survival Rate in 120pts









# 24-month Primary Patency in 120 pts





# 24-month Freedom from TLR in 120 pts








# 24-month Rutherford evolution – 120 pts










# 12m REFLOW results in perspective (lesions >20cm)






DCB: REFLOW results

BMS: Durability 200 study

# 24m REFLOW results in perspective (lesions > 20cm)





**DCB**: REFLOW results

# Conclusion

- Final 12- & 24-month results suggest that the LEGFLOW DCB is a valid and **effective** alternative to treat **"real-life"** long, complex and calcified femoropopliteal lesions
- With a 89.80% survival rate at 24-month, the LEGFLOW DCB proves it's safety